



La Science à l'œuvre pour le
at work for Canada

NRC Publications Archive Archives des publications du CNRC

An Integrated Patient Positioning and Beam Gating System for Radiation Therapy

El-Hakim, Sabry; Gerig, L.; MacPherson, M.; Vo, D.; Crljenko, T.

NRC Publications Record / Notice d'Archives des publications de CNRC:

<http://nparc.cisti-icist.nrc-cnrc.gc.ca/npsi/ctrl?action=rtdoc&an=5763536&lang=en>

<http://nparc.cisti-icist.nrc-cnrc.gc.ca/npsi/ctrl?action=rtdoc&an=5763536&lang=fr>

Access and use of this website and the material on it are subject to the Terms and Conditions set forth at

http://nparc.cisti-icist.nrc-cnrc.gc.ca/npsi/jsp/nparc_cp.jsp?lang=en

READ THESE TERMS AND CONDITIONS CAREFULLY BEFORE USING THIS WEBSITE.

L'accès à ce site Web et l'utilisation de son contenu sont assujettis aux conditions présentées dans le site

http://nparc.cisti-icist.nrc-cnrc.gc.ca/npsi/jsp/nparc_cp.jsp?lang=fr

LISEZ CES CONDITIONS ATTENTIVEMENT AVANT D'UTILISER CE SITE WEB.

Contact us / Contactez nous: nparc.cisti@nrc-cnrc.gc.ca.



National Research
Council Canada

Conseil national
de recherches Canada

Canada



National Research
Council Canada

Conseil national
de recherches Canada

Institute for
Information Technology

Institut de technologie
de l'information

NRC-CNRC

*An Integrated Patient Positioning and Beam Gating System for Radiation Therapy**

El-Hakim, S., Gerig, L., MacPherson, M., Vo, D., and Crljenko, T.
September 2002

* published Proceedings of the 21st Annual Meeting of the European Society for Therapeutic Radiology and Oncology (ESTRO 2002), September 15-16, 2002, Vol. 64, Supl. 1, Page S108, Prague, Czech Republic . NRC 44993.

Copyright 2002 by
National Research Council of Canada

Permission is granted to quote short excerpts and to reproduce figures and tables from this report, provided that the source of such material is fully acknowledged.

AN INTEGRATED PATIENT POSITIONING AND BEAM GATING SYSTEM FOR RADIATION THERAPY

L.H.Gerig, M. Macpherson, S. El-Hakim, D. Vo, T. Crljenko

Ottawa Regional Cancer Centre, Ottawa Canada
National Research Council of Canada, Ottawa, Canada
Carleton University, Ottawa, Canada

In radiation therapy it is recognized that margins are required around the CTV to ensure adequate minimal dose to the target tissues (ICRU 50). The margin between the PTV and CTV is a direct function of intra and inter fraction target and patient motion as well as daily setup deviation. In this work we describe a machine-vision respiratory-gating system, based on our reported photogrammetric patient position measurement technique¹. This system offers a means of providing digital direction for automatic patient setup as well as a real time gating signal to mitigate against patient motion during each fraction. The system comprises a minimum of two CCD cameras, each permanently mounted approximately two meters from the machine isocenter. Inherent contrast of retro-reflective markers on the patient's skin is achieved through illumination by infra-red lasers combined with infra-red band pass filters on the cameras. Any two cameras will then provide a stereo view of the markers and provide a basis for triangulation. The system includes a mechanism for automatic calibration and co-ordinate transformation into the linac frame (IEC) of reference providing a temporally invariant metric for patient position. Emphasis on image processing (e.g. image deformation and corrections for radial and decentering lens distortion) allows absolute marker positions to be reported in the linac frame of reference with an accuracy of +/- 0.5 mm, a precision of 0.2 mm and a temporal resolution of better than 150 ms. Integrating these functions provides a system that is simultaneously capable of monitoring mean patient position and real time variations due to patient motion and respiration. Duplicate systems in the simulator and treatment room allow direct referencing of patient treatment position to simulator films. The system in the simulator also allows simultaneous video capture of fluoroscopy video, time-synchronized to the vector position of radio-opaque reflective markers on the patient surface. The clinically relevant parameter is target motion in a plane perpendicular to the beam. Our system provides a tool to set the gating thresholds for each treatment field mapped from spatial boundaries determined in this plane (the beam's eye view). Performance of the system will be reviewed, and new applications will be discussed.

¹ LH Gerig, et al. "The Development and Clinical Application of a Patient Position Monitoring System." Videometrics III. Proc. SPIE 2350, Nov. 1994.