NRC Publications Archive Archives des publications du CNRC

Analysis of paralytic shellfish toxins using high field asymmetric waveform ion mobility spectrometry with liquid chromatography-mass spectrometry

Beach, Daniel G.; Melanson, Jeremy E.; Purves, Randy W.

This publication could be one of several versions: author's original, accepted manuscript or the publisher's version. / La version de cette publication peut être l'une des suivantes : la version prépublication de l'auteur, la version acceptée du manuscrit ou la version de l'éditeur.

For the publisher's version, please access the DOI link below. / Pour consulter la version de l'éditeur, utilisez le lien DOI ci-dessous.

Publisher's version / Version de l'éditeur:

<https://doi.org/10.1007/s00216-015-8488-6>

Analytical and Bioanalytical Chemistry, 407, pp. 2473-2484, 2015-01-27

NRC Publications Archive Record / Notice des Archives des publications du CNRC :

<https://nrc-publications.canada.ca/eng/view/object/?id=1c208da7-3df9-46a5-a3f1-d8de34813644>

<https://publications-cnrc.canada.ca/fra/voir/objet/?id=1c208da7-3df9-46a5-a3f1-d8de34813644>

Access and use of this website and the material on it are subject to the Terms and Conditions set forth at

<https://nrc-publications.canada.ca/eng/copyright>

READ THESE TERMS AND CONDITIONS CAREFULLY BEFORE USING THIS WEBSITE.

L'accès à ce site Web et l'utilisation de son contenu sont assujettis aux conditions présentées dans le site

<https://publications-cnrc.canada.ca/fra/droits>

LISEZ CES CONDITIONS ATTENTIVEMENT AVANT D'UTILISER CE SITE WEB.

Questions? Contact the NRC Publications Archive team at

PublicationsArchive-ArchivesPublications@nrc-cnrc.gc.ca. If you wish to email the authors directly, please see the first page of the publication for their contact information.

Vous avez des questions? Nous pouvons vous aider. Pour communiquer directement avec un auteur, consultez la première page de la revue dans laquelle son article a été publié afin de trouver ses coordonnées. Si vous n'arrivez pas à les repérer, communiquez avec nous à PublicationsArchive-ArchivesPublications@nrc-cnrc.gc.ca.

**Analysis of Paralytic Shellfish Toxins Using High Field Asymmetric Waveform Ion
Mobility Spectrometry with Liquid Chromatography – Mass Spectrometry**

Daniel G. Beach*¹, Jeremy E. Melanson², Randy W. Purves³

¹ National Research Council Canada, Measurement Science and Standards, Halifax, NS, Canada

² National Research Council Canada, Measurement Science and Standards, Ottawa, ON, Canada

³ University of Saskatchewan, Plant Sciences Department, Saskatoon, SK, Canada

*Corresponding author:

Dr. Daniel Beach

Tel.: +1 (902) 426-8274; Fax: +1 (902) 426-9413

E-mail address: daniel.beach@nrc-cnrc.gc.ca

National Research Council of Canada

Measurement Science and Standards

1411 Oxford Street

Halifax, NS, B3H 3Z1, Canada

Abstract

The analysis of paralytic shellfish toxins (PSTs) by LC-MS remains a challenge because of their high polarity, large number of analogues, and complex matrix in which they occur. Here we investigate the potential utility of high field asymmetric waveform ion mobility spectrometry (FAIMS) as a gas phase ion separation tool for analysis of PSTs by mass spectrometry. We investigate the separation of PSTs using FAIMS with two divergent goals; using FAIMS as a primary separation tool for rapid screening by ESI-FAIMS-MS or combined with LC in a multidimensional LC-ESI-FAIMS-MS separation. First, a survey of the parameters that affect the sensitivity and selectivity of PST analysis by FAIMS was carried out using ESI-FAIMS-MS. In particular, the use of acetonitrile as a gas additive in the carrier gas flow offered very acceptable separation of all PST epimeric pairs. A second set of FAIMS conditions was also identified, which focussed PSTs to a relatively narrow CV range allowing development of an LC-ESI-FAIMS-MS method for analysis of PST toxins in a complex mussel tissue extracts. The quantitative capabilities of this method were evaluated by analysis of a PST containing mussel tissue matrix material. Results compared favourably with analysis by the established LC- post column oxidation- fluorescence methodology with recoveries ranging from 70% to 106%, although sensitivity was somewhat reduced. The current work shows the promise of FAIMS as a tool for analysis of algal biotoxins in complex samples and outlines some critical requirements for its future improvement.

Keywords: Differential Mobility Spectrometry (DMS), Paralytic Shellfish Poisoning (PSP), Algal Toxins

38 **Introduction**

39 Paralytic shellfish toxins (PSTs) are potent neurotoxins produced by several marine
40 dinoflagellates and freshwater cyanobacteria that are bio-accumulated by bivalves and lead to
41 cases of paralytic shellfish poisoning (PSP) worldwide [1, 2]. The parent toxin of the class,
42 saxitoxin, is a Schedule 1 chemical weapon and has many structural analogues, which have been
43 identified in algae and shellfish (Figure 1). In most regulatory laboratories, the AOAC mouse
44 bioassay (MBA) has been used for routine analysis [3], but this method suffers from low
45 sensitivity and poor reproducibility [4]. Recently, the replacement of mouse bioassays in Europe
46 for the monitoring of lipophilic toxins [5] has led to with the adoption/use of instrumental
47 analytical methods, capable of more sensitive and reliable quantification as well as confirmation
48 of the chemical nature of the detected toxins. The most commonly used chemical analytical
49 methods for PSTs are those based on liquid chromatography (LC) coupled with fluorescence
50 detection and using pre- or post- column oxidation reactions producing fluorescent PST
51 derivatives [6-8]. These methods have some drawbacks related to system upkeep and lack of
52 absolute structural confirmation, which is only possible using mass spectrometry detection. In
53 general, shellfish toxin analysis over the last several years has been moving towards the use of
54 liquid chromatography coupled to tandem mass spectrometry detection (LC-MS/MS) with
55 electrospray ionization (ESI), which has been successfully implemented for the regulatory
56 monitoring of lipophilic shellfish toxins [9-11].

R ₁	R ₂	R ₃	R ₄	Toxin ^a	[M+H] ⁺ , [M+H-SO ₃] ⁺
H	H	H		STX	300
H	H	OSO ₃ ⁻		GTX2	396, 316
H	OSO ₃ ⁻	H	—O—C(=O)—NH ₂	GTX3	396, 316
OH	H	H		NEO	316
OH	H	OSO ₃ ⁻		GTX1	412, 332
OH	OSO ₃ ⁻	H		GTX4	412, 332
H	H	OSO ₃ ⁻	—OH	dcGTX2	353, 273
H	OSO ₃ ⁻	H		dcGTX3	353, 273

^a STX = saxitoxin, NEO = neosaxitoxin, GTX = gonyautoxins, dc = decarbamoyl

Fig. 1 structures of PSTs analysed

Compared with lipophilic classes of algal toxins, the analysis of PSTs by LC-MS remains a significant instrumental challenge for several reasons including (a) their highly polar character limiting their retention in reverse phase LC, (b) the large number of structurally similar analogues including several epimeric pairs (Fig. 1), (c) labile precursor ions whose in-source fragments share *m/z* values with precursors or product ions of other analogues and (d) chemical interference and ionization suppression resulting from the complex shellfish tissue extract matrix. Despite these challenges, a number of LC-MS methods have been reported for the analysis of PSTs using hydrophilic interaction liquid chromatography (HILIC) mode of LC separation, which allows for retention and separation of highly polar PSTs [12-15]. This previous work has demonstrated the potential of the technique as well as some of its current limitations, including the presence of ESI signal suppression and matrix interferences [12].

High Field Asymmetric Waveform Ion Mobility Spectrometry (FAIMS), also known as Differential Mobility Spectrometry (DMS), is a relatively new mode of analytical separation that was developed for use with mass spectrometry by Guevremont *et al.* [16]. Since that time, several devices based on the same principles of separation have been made commercially

available by MS instrument manufacturers [17-20]. The FAIMS device acts as a continuous ion filter to separate ions in the gas phase, at atmospheric pressure after they are produced by ESI but prior to entering the MS. The mechanism of separation in FAIMS is based on differences in the mobility characteristics of ions as they are subjected to an electric field oscillating between high and low field strengths [17, 21-24]. The main advantage of using FAIMS in analytical mass spectrometry lies in the orthogonality of the separation (based on the size, charge, shape and polarisability of an ion), to the separations in both liquid chromatography and mass spectrometry. This orthogonality results in increased selectivity for isolation of a targeted analyte from interfering species such as other isobaric analytes or matrix components. Thus, analytical MS methods incorporating a FAIMS filter step prior to MS analysis are able to selectively resolve and quantify species that otherwise cannot be selectively analysed by ESI-MS or LC-ESI-MS alone. When used as a primary separation tool, ESI-FAIMS-MS can provide sufficient selectivity to perform very rapid analyses (ca. 30 seconds/sample) of samples without the need for LC separation [24, 25]. For more complex samples that cannot be analysed by direct infusion electrospray, FAIMS can be combined with LC (LC-ESI-FAIMS-MS) to perform a highly selective multidimensional separation [22].

The current work aims to evaluate FAIMS as a possible means of improving selectivity in LC-MS analysis of PSTs, where interference from matrix compounds has been identified as a limitation of current LC-MS methodology [12]. We begin by investigating the FAIMS separation of PST isomers from one another by direct ESI-FAIMS-MS, with the secondary goal of focusing PSTs to similar CV values suitable for LC-ESI-FAIMS-MS. Recent literature has highlighted the significant benefits that use of polar solvent gas modifiers can impart to differential mobility separations [17, 20, 26, 27]. Here we investigate the impact of acetonitrile vapour on FAIMS

separation of PSTs. An LC-ESI-FAIMS-MS method was then developed for analysis of PSTs in mussel tissue extracts. Finally, in order to evaluate the quantitative capabilities of this technique, a PST mussel tissue reference material was analysed and results were compared with those obtained using LC-FLD with post-column oxidation.

Experimental

Optima LC-MS grade Acetonitrile (CH_3CN) was purchased from Fisher (Ottawa, ON, Canada). Ammonium Formate (98%) was purchased from Fisher Chemicals (Fairlawn, NJ, USA). Formic Acid (98%) was purchased from EMD (Darmstadt, Germany). Deionized water (DIW) was produced by passing distilled water through a Milli-Q Gradient A10 de-ionized water system (Millipore, USA).

All PST standards and matrix materials were provided by the National Research Council Canada (Halifax, NS, Canada). These included the certified reference material calibration solutions for the toxins in Fig. 1: CRM-STX-f, CRM-NEO-c, CRM-GTX2&3-c, CRM-GTX1&4-c and CRM-dcGTX2&3-b. Tissue matrix materials analyzed included CRM-Zero-Mus, a mussel tissue homogenate certified to be free of PSTs as well as two different pilot scale mussel tissue reference materials [28] whose PST concentrations, homogeneity and stability have been well characterized by an established analytical method, LC-pCox-FLD [8]. Extractions of the blended mussel tissues were prepared by using a modified version of a previously published AOAC extraction procedure [6]. Briefly, 4 g of mussel tissue homogenate were mixed with 4 mL of 0.1 M HCl and boiled for 10 min. After centrifugation, the supernatant was passed through a 60 mg OASIS HLB solid phase extraction cartridge (Waters, Milford, MA), de-proteinated by adding one volume of acetonitrile and passed through a 4.5 μm pore size filter. Quantification of PSTs in tissue reference material was carried out by spiking 100 μL of deionized water or of

120 mixed toxin standard into separate 400 μ L aliquots of extract to construct a three point standard
121 addition calibration curve used to determine the concentration of the PSTs in the tissue.

122 *ESI-FAIMS-MS and LC-ESI-FAIMS-MS - System 1*

123 A Thermo Fisher LTQ-Orbitrap (Bremen, Germany) equipped with a Thermo Fisher FAIMS
124 system was used for both infusion and LC-based analysis. ESI-FAIMS-MS was carried out by
125 infusing toxin CRMs diluted to approximately 1 - 10 μ M in 1:1 Acetonitrile:Water with 0.1%
126 acetic acid at 5 μ L/min. MS source parameters for these low-flow infusions included a source
127 temperature of 100 $^{\circ}$ C, capillary temp of 100 $^{\circ}$ C, capillary voltage of 10 V, a tube lens voltage of
128 40 V, spray voltage of + 3500 V, sheath gas of 10, auxiliary gas of 5 and sweep gas of 0 (all
129 arbitrary units). Mass spectral data was collected in full scan mode from m/z 250 - 450 or in
130 selected ion monitoring mode, all with an IT fill time of 100 ms unless otherwise noted. FAIMS
131 outer bias voltage was 10 V and gas flows, dispersion voltage and electrode temperatures are
132 cited for individual experiments throughout. For flow injection analysis and LC-ESI-FAIMS-MS
133 experiments carried out at a flow rate of 200 μ L/min, harsher source parameters were used to
134 ensure better desolvation prior to introduction into FAIMS. These included a source temperature
135 of 300 $^{\circ}$ C, capillary temp of 350 $^{\circ}$ C, capillary voltage of 10 V, a tube lens voltage of 40 V, spray
136 voltage of + 3500 V, sheath gas of 35, auxiliary gas of 30 and sweep gas of 0 (all arbitrary units).

137 Liquid chromatography was performed using an Agilent 1200 LC system with an Acquity UPLC
138 1.7 μ m BEH Amide 2.1 x 100 mm column (Waters, Milford, MA, USA) operated in HILIC
139 mode, similar to conditions reported previously [14]. LC flow rate was 250 μ L/min and injection
140 volume was 10 μ L. Mobile phase used was 2 mM ammonium formate pH 3.5 (A) and 0.1%
141 formic acid in acetonitrile (B) using a linear gradient from 75% to 55% B in 10 min, held for 2
142 min, before returning to the initial conditions.

System 2: ESI-FAIMS-MS Investigation of gas additives

A FAIMS-equipped Thermo Fisher TSQ Quantum (San Jose, CA) was used for ESI-FAIMS-MS studies using gas modifiers. The standard setup was modified as described previously to enable the use of gas additives [20]. Briefly, a Waters UPLC system was used to deliver a reproducible flow of solvent into a mixing T installed in-line with the buffer gas flow being delivered to the FAIMS electrodes. System 2 also used a modified electrode set having a 15 mm inner electrode and an improved desolvation region [18]. This resulted in a narrower electrode gap and overall higher electric field strength at a given applied dispersion voltage than system 1, which used a 13 mm inner electrode. A Harvard apparatus syringe pump was used to deliver PST standard solutions to the ESI needle tip at a flow rate of 15 μ L/min. The ESI needle was operated in positive ion mode and MS detection was carried out in selected reaction monitoring (SRM) scan mode using the conditions shown in Table 1. Optimized conditions for analysis of the mixture of PSTs in Fig. 1 included a dispersion voltage of 4400 V, a gas flow of 4 L/min N₂ with 1.25% acetonitrile as the modifier and inner/outer electrode temperatures of 35/55 °C and CV values and SRM conditions in Table 1. Alternative FAIMS parameters for separation of specific epimeric pairs are described in figure captions and text in Results and Discussion.

Table 1: FAIMS and MS/MS conditions used for the separation of PSTs using 1.25% acetonitrile as a carrier gas modifier on System 2.

Analyte	Precursor Ion m/z	Product Ion m/z	Collision Energy	Compensation Voltage
STX	300	204	21	-70
NEO	316	298	15	-70
GTX1	412	332	15	-38
GTX2	396	316	7	-36
GTX3	396	298	15	-33
GTX4	412	314	15	-30
dcGTX2	353	273	7	-33
dcGTX3	353	255	15	-37

Results and Discussion

Depending on the analytical method being developed, FAIMS optimization for PSTs was carried out towards one of two divergent goals. For selective direct analysis of PSTs by ESI-FAIMS-MS, the epimeric pairs GTX2/GTX3, GTX1/GTX4, and dcGTX2/dcGTX3 (Figure 1) must be separated from one another since they are not resolved in mass spectrometry. Further, the abundant product ions and in-source fragments of GTX2 and GTX3 are sulphate loss product ions identical in composition to protonated NEO (Figure 1), requiring complete separation of these compounds prior to MS analysis. On the other hand, in LC-ESI-FAIMS-MS, isolation of PSTs from interfering matrix components rather than from one another is the primary goal. The LC separation of PST epimers from one another is possible because of significant differences in molecular conformation [12, 29]. Since FAIMS acts as an ion filter, the number of simultaneous CV values that must be monitored at any one time during an LC separation must be minimized in order to minimize the impact on the overall duty cycle of the analysis. Thus, it becomes desirable to focus co-eluting analytes to the same CV values away from interfering matrix components, which have previously been identified as a limitation to LC-MS analysis of PSTs [12].

178 *Separation of PSTs by FAIMS on System 1*

179 The FAIMS separation properties of PSTs were investigated by infusing toxin standards at low
180 flow rates for direct ESI-FAIMS-MS analysis. Each FAIMS parameter was investigated by
181 scanning compensation voltage (CV), initially from -50 V to 50 V. The parameters that were
182 investigated included dispersion voltage (DV), carrier gas composition and flow rate, and
183 electrode temperature. Depending on the differential mobility behaviour of the analytes they can
184 be more effectively transmitted and/or separated through FAIMS with either positive or negative
185 dispersion voltages, referred to as mode P2 and P1, respectively [30]. Under all sets of conditions
186 examined on system 1, PST transmission was around 30-fold higher in P2 mode than in P1
187 mode, limiting the practical utility of the latter for PST separation. In mode P2, an exponential
188 decline in sensitivity was observed below the maximum operating value for dispersion voltage of
189 - 5000 V, which was then used for all subsequent analyses.

190 Increased electrode temperature had the benefit of improving transmission of the un-sulphated
191 PSTs, STX and NEO but the un-desirable effect of increasing in-source fragmentation of the
192 labile GTXs. A decrease in electrode temperature also had a favourable impact on separation
193 between GTX epimeric pairs, as can be seen in the selected ion compensation voltage (SICV)
194 spectra in Figure 2. The behaviour of α - and β - sulphated PST epimers during collisional induced
195 dissociation has been well studied and these compounds differ greatly in lability both in the MS
196 source and during MS/MS experiments [12, 29]. In Figure 2, SICV spectra of the m/z 396
197 $[M+H]^+$ ion and its m/z 316 sulphate loss source fragment can be used as an approximation for
198 the less labile GTX3 and the more labile GTX2, respectively. At high values of 70/90 °C
199 inner/outer electrode temperature, little separation was observed between the epimers but better
200 separation is observed at the lowest value of 35/45 °C.

Figure 2: Extracted ion compensation voltage spectra showing the effect of electrode temperature on GTX2 (m/z 316) and GTX3 (m/z 396) separation at (A) 70/90 °C, (B) 50/70 °C and (C) 35/45 °C inner/outer electrode temperatures.

At sample infusion flow rates, the FAIMS carrier gas flow rate did not have a significant impact on either sensitivity or epimer resolution. A gas flow rate of 3.5 L/min gave improved sensitivity and separation compared with lower flow rates, while higher flow rates resulted in instability of the ion signal. Gas composition proved to have a significant impact on sensitivity and CV of transmission for PSTs, but less impact on separation of epimeric pairs. In P2 mode, the addition of He to the N₂ buffer gas resulted in a significant increase in signal intensity as shown in Figure 3 for STX, and a shift to more negative CV values up to 50% He, which is the

maximum the instrument allows due to the risk of electrical discharge. An alternative carrier gas modifier, CO₂, has previously been reported to provide more gentle conditions for the detection of fragile ions as well as superior separation for small polar analytes [24, 25]. For PSTs, the addition of CO₂ resulted in a significant loss of overall intensity in P2 mode without any improvement in epimer separation of epimeric pairs. In P1 mode, where CO₂ has previously been shown to be the most beneficial, addition of CO₂ gave a modest improvement in sensitivity, but without enhancement of epimer separation. The relative intensity of the labile [M+H]⁺ ion of GTX 1 at *m/z* 412 did however increase with the addition of CO₂ to the point where it was nearly equal in intensity with the sulphate loss in-source fragment ion at *m/z* 332, supporting a gentler introduction into the mass spectrometer than without CO₂. However, compared with 50 % He in N₂ as buffer gasses, the absolute intensity was significantly lower when CO₂ was added.

Figure 3: Extracted ion compensation voltage spectra showing the impact of increasing % He in N₂ for the analysis of protonated saxitoxin at *m/z* 300 by ESI-FAIMS-MS.

The ion filtering effect of FAIMS significantly reduces the total ion current introduced into the mass spectrometer at any one time. This has a beneficial impact for ion trap mass analysers, which can be operated at much higher fill times than when all ions produced by ESI

are introduced into MS [31]. For PST toxin standards at 5 μ M, ion trap fill times could be increased from the default value of 10 msec to above 1000 msec without trap saturation. Provided the IT fill time is set to below the time spent at each CV step, it is possible to increase S/N in ESI-FAIMS-MS at no cost to analysis time or FAIMS resolution. Greater care must be taken in LC-ESI-FAIMS-MS experiments where achieving an acceptable number of data points across an LC peak becomes an important consideration.

ESI-FAIMS-MS/MS Analysis of PSTs Using Gas Additives on System 2

The recent FAIMS and DMS literature has been dominated by the use of polar solvent vapours as dopants in low % amounts in the buffer gas to enhance selectivity of the FAIMS separation [17, 20, 26, 27]. We investigated this possibility for PSTs using a second ESI-FAIMS-MS system (System 2), which was modified to introduce solvent into the buffer gas flow. As reported previously for a broad range of other analytes [20], acetonitrile was found to have a favourable impact on both the sensitivity of PST analysis and on epimer separation in FAIMS. Without the use of acetonitrile, GTX1 and GTX4 showed no separation using 100 % N_2 and were detected in P2 mode ($DV = -4000$ V) on system 2, as shown in trace (a) of Figure 4A. When 0.6 % acetonitrile was added to the carrier gas (trace (b) of Figure 4A) baseline separation of the two epimers was observed in P1 mode ($DV = +4000$ V). The broad peak at $CV = -41$ V can be attributed to species transmitted through FAIMS as dimers or multimers at different CV values than the $[M + H]^+$ ions detected at $CV = -32$ V (GTX1) and -27 V (GTX4), but which then decompose to monomers upon introduction into mass spectrometry. This assignment is supported by the fact that the $CV = -41$ V peak increases to 80% relative intensity at higher analyte concentrations (8 μ M). Using the more sensitive mass spectrometer in single reaction

253 monitoring (SRM) mode in system 2, toxin standards were analysed at lower analyte
254 concentrations (as low as 0.1 μ M) where the formation of multimeric species was minimized.
255 The baseline separation of GTX1 and GTX4 epimers achieved using ESI-FAIMS-MS with 0.6%
256 acetonitrile as a gas additive at a dispersion voltage (DV) of 4000 V allowed for full scan MS
257 spectra of each toxin to be acquired selectively without the need for liquid separation as shown in
258 Figure 4B and 4C. Since the behaviour of PSTs during collision induced dissociation has been
259 well characterized [12, 29], these spectra serve as additional evidence that the CV peaks in Fig. 4
260 represent the true separation of GTX1 and GTX4 epimers rather than of different gas phase
261 conformations of a mixture of isomers resulting from solvation or protonation.

Fig. 4: The effect of acetonitrile vapour dopant on the FAIMS separation of GTX1 and GTX4 using system 2. (A) 100% N₂ at DV = -4000 V is shown in trace (a) and 0.6% acetonitrile in N₂ at DV = 4000 V is shown in trace (b) using SRM parameters in Table 1. Full scan MS spectra collected using 0.6% acetonitrile at CV = -32 V for GTX1 (B) and at CV = -26 V for GTX4 (C).

The temperature gradient between the inner and outer FAIMS electrodes was found to be a critical variable in the separation of PST epimers. The optimal 20 °C gradient for separation of

270 GTX epimers used an inner/outer electrode temperature of 35/55 °C. When this gradient was
271 decreased, the separation deteriorated. For example, GTX2 and GTX3, which were completely
272 separated at 20 °C were only partially separated at 10 °C, and with equal electrode temperatures
273 showed only minimal separation. However, the improved separation using an electrode
274 temperature gradient was achieved at the cost of sensitivity of analysis of the sulphated PSTs
275 with a decrease of between 2 and 10 fold, depending on the analogue. It was therefore not
276 possible to achieve optimal separation and sensitivity for all PSTs under a single set of FAIMS
277 conditions. Conditions optimized for separation of one pair of GTX epimers were not completely
278 optimal for another. In general, improved GTX separation at increased DV and % acetonitrile
279 dopant gave a broadening of the STX peak and pushed it to higher CV values. Of the epimer
280 pairs, GTX 2 and GTX 3 were the most difficult to separate from one another. Figure 5 shows all
281 PSTs analysed at DV = 4400 V with 1.25 % acetonitrile additive, optimal conditions for GTX2
282 and GTX3 separation, which gave acceptable separation of all epimeric pairs. Under the
283 conditions described in Table 1, all GTX epimer pairs are separated with a resolution of at least
284 0.6 (GTX2 and GTX3) and baseline resolution for other pairs. Minor peaks observed between
285 $\sim CV = -45$ V and -60 V can be attributed to multimeric species of PST ions, which under these
286 conditions are minimized compared with the protonated species. The more gentle conditions for
287 introduction of labile GTX ions into MS afforded by the use of gas additives allowed for all
288 species to be detected as their $[M+H]^+$ ions in SRM scan mode.

290

291 **Figure 5:** Analysis of a PST standard mixture using 1.25% acetonitrile as an additive in N₂ gas at
 292 a dispersion voltage of 4400 V with an inner/outer electrode temperature of 35/55 °C on System
 293 2. SRM detection conditions for each compound are given in Table 1.

294

295 *FAIMS Optimization for LC-ESI-FAIMS-MS on System 1*

296 It is evident that for selective analysis of PSTs by direct ESI-FAIMS-MS(/MS) the use of
 297 gas additives is highly desirable as baseline separation between epimers was not otherwise
 298 achieved. However, other more significant challenges related to ionization exist that currently
 299 limit the possibility of doing direct analysis of PSTs in complex environmental samples by this
 300 technique. In particular, the complex shellfish tissue extracts in which PSTs must typically be
 301 quantified cannot currently be analysed by direct electrospray ionization due to severe ionization
 302 suppression from matrix components. Analysis of PSTs in these complex samples must instead

be done using LC-ESI-FAIMS-MS. In this instrumental configuration, separation of PST epimers is not required as these can readily be separated by LC. However, the CV switching time of FAIMS limits the number of different CV values that can be simultaneously monitored at any point in a chromatographic run. It therefore became desirable to focus PSTs to a narrower CV space, but separated from potential matrix interference. The need to monitor multiple CVs simultaneously during an LC run also has a significant negative impact on analytical sensitivity and limits of detection. For this purpose, the use of 50% He in N₂ carrier gas without solvent vapour dopant was favoured as it gave the best sensitivity observed on system 1 while also transmitting the examined PSTs at a smaller range of CV values than with dopant on system 2.

In order to scale-up the MS source and FAIMS conditions from the 5 μ L/min flow rate used for infusion to the 200 μ L/min flow rate of the LC method, several parameters needed to be re-optimized. This was done by carrying out CV scans in single-ion monitoring mode for [M+H]⁺ and [M+H-SO₃]⁺ ions using flow injection analysis of toxin standards. Initially, source parameters including spray voltage, temperature and capillary voltage needed to be optimized in order to achieve acceptable desolvation at the higher flow rate, but FAIMS parameters were observed to be dependent on flow rate and source parameters as well. In particular, the use of higher inner/outer electrode temperature (90/105 °C) resulted in roughly a 4-fold increase in sensitivity at the higher flow rate. This temperature change caused an increase in-source fragmentation to the point where all sulphated toxins were much more sensitively detected as their [M+H-SO₃]⁺ source fragments than as [M+H]⁺ ions. Under these optimized high-flow conditions, little separation was observed between PST epimers as shown in Figure 6. This has the added benefit of reducing the number of CVs that need to be monitored in an LC-ESI-FAIMS-MS experiment, which is important for sensitivity of the analysis. The additional peak

detected in the dcGTX2/dcGTX3 standard was ruled out as being separated dcGTX3 by LC-ESI-FAIMS-MS experiment, which showed that no toxin was detected at CV = - 22.5 V. This peak can instead be assigned as a buffer or system contaminant which is effectively filtered by FAIMS during analysis of dcGTX2 and dcGTX3 at CV = - 15.6 V.

Fig. 6: CV Optimization for transmission of PSTs standards through FAIMS on system 1 at LC flow rate (200 μ L/min) using flow injection analysis. STX, NEO, GTX3 and GTX4 were detected as their $[M+H]^+$ ions at m/z 300, 316, 396 and 412, respectively while GTX1, GTX2, dcGTX2 and dcGTX3 were detected as their $[M+H-SO_3]^+$ source fragments at m/z 332, 316, 273 and 273, respectively, all in single ion monitoring mode. * labelled peak at - 21 V in dcGTX2/3 trace is a system contaminant separated from dcGTXs by FAIMS.

The limiting factor for analytical sensitivity is the duty cycle of the FAIMS device, which has a switching time between different CV values of about 100 ms, the time required to empty the device of ions that experienced a particular CV. This means that limiting the number of CV values that must be monitored simultaneously is an effective way of limiting any sensitivity losses observed when using FAIMS in combination with LC. Two approaches were investigated for limiting the number of monitored CVs. These included using time periods with a limited number of optimized CV values at a given retention time or reducing the number of monitored

CVs to three values which provided coverage of all analytes close to, but not at their CV maxima. Table 2 shows optimized CV values for each analyte under the high flow conditions described in the previous sections, as well as their retention time in HILIC-LC. There are three distinct groupings of PSTs by retention time in the HILIC separation, (a) GTX1 and GTX2 prior to 5 min, (b) GTX3, GTX4 and dcGTX3 between 5 and 6.3 min and (c) STX and NEO after 6.3 minutes, while dcGTX2 and dcGTX3 span the first two periods eluting between 4.8 and 5.2 min. It was therefore possible to create a method where a maximum of 3 CV values were monitored at any one time as seen for PST standards in Fig. 7A and PSTs in a mussel tissue reference material in Fig. 7B. The relatively poor detection of GTX4 in Fig. 7B is due primarily to the lower concentration of this analyte in the reference material (0.2 μ M in the analysed extract). The primary limitation of this time-binned CV method was that small drifts in retention time over the course of an LC sequence were problematic, particularly the change at 6.3 min which either risked moving onto the front of the STX peak or the tail of the dcGTX3 peak.

Fig 7: LC-ESI-FAIMS-MS analysis of PSTs (A) in a mixed standard and (B) in mussel tissue extracts using retention time bins with optimized CV values for each analyte (details in text) in single ion monitoring mode.

In order to evaluate the quantitative capabilities of LC-ESI-FAIMS-MS for the analysis of PSTs in mussel tissue a more robust method needed to be developed that was not as sensitive to small variations in retention time, which are not uncommon in the analysis of tissue extracts by LC. For this purpose, close approximations to the maximum CV values of each toxin was used to limit the number of CV values which needed to be monitored to 3, but each of these was monitored continuously over the course of the analysis. These individual channels included GTX1 and GTX4 and NEO detected at CV= - 17 V, dcGTX2 and dcGTX3 detected at CV = - 14.5 and STX detected at CV = - 19 V. This more robust method was less sensitive due to the fact that toxins were not detected at their optimum CV values, but was more suitable for analysis of large sample sets where retention time could be expected to drift slightly.

Table 2: FAIMS, LC and MS parameters used to detect PSTs in a mussel tissue reference material as well as values determined by LC-ESI-FAIMS-MS and those determined using an established technique, LC-pcox-FLD

PSP-Mus Pilot	CV (V)	r.t. (min)	[M+H] ⁺ (m/z)	m/z monitored by LC-ESI-FAIMS-MS	LC-pcox-FLD (μmol/kg tissue ± SD) ^a	LC-ESI-FAIMS-MS (μmol /kg tissue ± SD ^b)
GTX-1	- 17.7	4.7	412	332 ^c	2.7 ± 0.3	1.9 ± 0.3
GTX-2	- 17.8	4.6	396	316 ^c	5.6 ± 0.6	4.3 ± 0.1
GTX-3	- 17.8	5.4	396	316 ^c	1.6 ± 0.2	1.7 ± 0.2
GTX-4	- 19.7	5.7	412	332 ^c	0.69 ± 0.07	< LOQ
dcGTX-2	- 15.8	4.9	335	273 ^c	8.3 ± 0.8	6.5 ± 0.3
dcGTX-3	- 15.8	5.9	335	273 ^c	2.2 ± 0.2	< LOQ
STX	- 19.1	7.0	300	300	2.6 ± 0.3	2.68 ± 0.07
NEO	- 17.5	7.2	316	316	1.4 ± 0.2	1.5 ± 0.1

^a Results from an established analytical technique, LC-post column oxidation - fluorescence [8] used previously to characterize the mussel tissue reference material [28].

^b Standard deviation of three replicate standard addition experiments

^c [M+H-SO₃]⁺ product ions formed as the result of in source fragmentation of protonated PSTs

The quantitative capabilities of LC-ESI-FAIMS-MS for the analysis of PSTs were evaluated by analysing a second mussel tissue reference material using a standard addition method and the continuous CV monitoring approach described above. This reference material consisted of a mixture of contaminated mussel tissue and toxin-containing algae previously produced on a pilot scale for the standardization of PST measurements [28]. For the most part, quantitative results from LC-ESI-FAIMS-MS agreed well with those from an established method for PST analysis, LC-post column oxidation-FLD [8]. The biggest limitation to these results is the relatively poor sensitivity of the current instrumental setup, which was not able to detect GTX4 or dcGTX3 in the reference material. Both the selectivity and sensitivity of the analysis could be improved by use of a triple quadrupole instrument in single reaction monitoring (SRM) scan mode, which itself would be much more suitable for quantitative analysis in complex tissue extracts.

Conclusions

Our work has outlined the two ways in which FAIMS separation can be used for the analysis of PSTs, either by separation of PST analogues for direct ESI-FAIMS-MS or by focussing of analytes to a narrow CV range away from matrix interference for multidimensional LC-ESI-FAIMS-MS/MS analysis. For direct analysis, the use of low % levels of acetonitrile (~ 1%) as an additive in the carrier gas was necessary to achieve adequate separation of all PST epimeric pairs not resolved by mass spectrometry. The significant positive impact of solvent vapour dopants is consistent with the recent literature in planar DMS devices [17, 27] as well as the cylindrical FAIMS device used here [20]. For LC-ESI-FAIMS-MS, 50 % He in N₂ without the use of gas additives provided the best balance of sensitivity while limiting the number of CV values which needed to be simultaneously monitored at a given LC retention time. Faster CV switching time

would allow for simultaneous monitoring of a larger number of CV values which would allow for the superior selectivity of the FAIMS device operated with gas additives to be exploited to a greater extent in the future. For improved analysis of highly labile compounds like PSTs by LC-ESI-FAIMS-MS, better desolvation without inducing in-source fragmentation than could be achieved here with System 1 would also be highly advantageous. This limitation was less evident with the alternative source design on System 2, and should have less of an impact going forward. These highlighted limitations represent practical aspects to the implementation of FAIMS technology rather than fundamental limitations of the technique, which we have shown here to be promising for the analysis of challenging classes of analytes, such as PSTs, in complex samples such as mussel tissue extracts. Future work will involve implementing LC-ESI-FAIMS-MS/MS with gas additives for improved selectivity and an investigation of direct PST quantification using ESI-FAIMS-MS/MS.

Acknowledgements

The authors gratefully acknowledge the technical assistance of Krista Thomas and Margaret McCooeye as well as the support and editorial assistance of Pearse McCarron and Michael Quilliam. The authors would like to thank Thermo Fisher for the loan of the TSQ Quantum.

422 **References**

- 423 1. Vale P (2014) Saxitoxin and Analogs: Ecobiology, Origin, Chemistry, and Detection. In:
424 Botana LM (ed) Seafood and Freshwater Toxins: Pharmacology Physiology and Detection, third
425 edition. CRC Press, Boca Raton, FL.
- 426 2. Quilliam MA (2003) The role of chromatography in the hunt for red tide toxins. *J.*
427 *Chromatogr. A.* 1000: 527-548
- 428 3. MBA, AOAC 959.08 Official Methods of Analysis (2000) 17th Ed., AOAC
429 INTERNATIONAL, Gaithersburg, MD, Method 959.08
- 430 4. Hess P, Grune B, Anderson DA, Aune T, Botana LM, Caricato P, van Egmond HP,
431 Halder M, Hall S, Lawrence JF, Moffat C, Poletti R, Richmond J, Rossini GP, Seamer C,
432 Vilageliu JS (2006) Three Rs Approaches in Marine Biotxin Testing. *Alt. Labor. Anim.* 34: 193-
433 224
- 434 5. Anonymous (2011) Commission Regulation (EU) No 15/2011 of 10 January 2011
435 amending Regulation (EC) No 2074/2005 as regards recognised testing methods for detecting
436 marine biotoxins in live bivalve molluscs. Official Journal of the European Union L 006 of 1
437 January 2011: 3-6
- 438 6. Anonymous (2005) AOAC Official method 2005.06 Quantitative determination of
439 paralytic shellfish poisoning toxins in shellfish using pre-chromatographic oxidation and liquid
440 chromatography with fluorescence detection. AOAC International, Gaithersburg.

- 441 7. Rourke WA, Murphy CJ, Pitcher G, van de Riet JM, Burns BG, Thomas, KM, Quilliam
442 MA (2008) Rapid postcolumn methodology for determination of paralytic shellfish toxins in
443 shellfish tissue. *J. AOAC Int.* 91: 589-597
- 444 8. Van de Riet JM, Gibbs RS, Chou FW, Muggah FW, Rourke WA, Burns G, Thomas G,
445 Quilliam MA (2009) Liquid Chromatographic Post-Column Oxidation Method for Analysis of
446 Paralytic Shellfish Toxins in Mussels, Clams, Scallops and Oysters: Single-Laboratory
447 Validation. *J. AOAC Int.* 92: 1690-1704
- 448 9. Brana-Magdalena A, Leao-Martins JM, Glauner T, Gago-Martinez A (2014)
449 Intralaboratory Validation of a Fast and Sensitive UHPLC/MS/MS Method with Fast Polarity
450 Switching for the Analysis of Lipophilic Shellfish Toxins. *J. AOAC Int.* 97: 285-292
- 451 10. McNabb P, Selwood AI, Holland PT, Aasen J, Aune T, Eaglesham G, Hess P, Igarishi M,
452 Quilliam M, Slattery D, Van de Riet J, Van Egmond H, Van den Top H, Yasumoto T (2005)
453 Multiresidue method for determination of algal toxins in shellfish: single-laboratory validation
454 and interlaboratory study. *J AOAC Int* 88 (3):761-772
- 455 11. van den Top HJ, Gerssen A, McCarron P, van Egmond HP (2011) Quantitative
456 determination of marine lipophilic toxins in mussels, oysters and cockles using liquid
457 chromatography-mass spectrometry: inter-laboratory validation study. *Food Additives &*
458 *Contaminants: Part A* 28 (12):1745-1757
- 459 12. Dell'Aversano C, Hess P, Quilliam MA (2005) Hydrophilic interaction liquid
460 chromatography-mass spectrometry for the analysis of paralytic shellfish poisoning (PSP) toxins.
461 *J. Chromatogr. A.* 1081: 190- 201

- 462 13. Zhou L, Yin Y, Fu W, Qui B, Lin, Yang Y, Zheng L, Li J, Chen G (2013) Determination
463 of paralytic shellfish poisoning toxins by HILIC-MS/MS coupled with dispersive solid phase
464 extraction. *Food Chem.*137: 115-121
- 465 14. Blay P, Hui JPM, Chang J, Melanson JE (2011) Screening for multiple classes of marine
466 biotoxins by liquid chromatography-high-resolution mass spectrometry. *Anal. Bioanal. Chem.*
467 400: 577-585
- 468 15. Watanabe R, Matsushima R, Harada T, Oikawa H., Murata M, Suzuki T (2013)
469 Quantitative determination of paralytic shellfish toxins in cultured toxic algae by LC-MS/MS.
470 *Food Addit. & Contam. A.* 8:1351-1357
- 471 16. Purves R, Guevremont R, Day S, Pipich CW, Matyjaszczyk S (1998) Mass spectrometric
472 characterization of a high-field asymmetric waveform ion mobility spectrometer. *Rev. Sci.*
473 *Instrum.* 69: 4094 - 4105
- 474 17. Schneider BB, Covey TR, Coy SL, Krylov EV, Nazarov EG (2010) Chemical Effects in
475 the Separation Process of a Differential Mobility/Mass Spectrometer System *Anal. Chem.* 82:
476 1867-1880
- 477 18. Dunyach J, Prasad S, Belford M (2014) Effect of Electrode Geometry on FAIMS Gas
478 Flow Focusing and Lateral Diffusion. Thermo Scientific poster note PN-64099-ASMS-EN-
479 0614S. ([http://www.thermoscientific.com/content/dam/tfs/ATG/CMD/cmd-documents/sci-](http://www.thermoscientific.com/content/dam/tfs/ATG/CMD/cmd-documents/sci-res/posters/ms/events/asms2014/PN-64099-Electrode-Geometry-ASMS2014-PN64099-EN.pdf)
480 [res/posters/ms/events/asms2014/PN-64099-Electrode-Geometry-ASMS2014-PN64099-EN.pdf](http://www.thermoscientific.com/content/dam/tfs/ATG/CMD/cmd-documents/sci-res/posters/ms/events/asms2014/PN-64099-Electrode-Geometry-ASMS2014-PN64099-EN.pdf),
481 accessed on Aug 1st, 2014)

- 482 19. Shvartsburg AA, Smith RD, Wilks A, Koehl A, Ruiz-Alonso D, Boyd B (2009) Ultrafast
483 Differential Ion Mobility Spectrometry at Extreme Electric Fields in Multichannel Microchips.
484 *Anal. Chem.* 81: 6489-6495
- 485 20. Purves RW, Ozog A, Ambrose SJ, Prasad S, Belford M, Dunyach J (2014) Using Gas
486 Modifiers to significantly improve sensitivity and selectivity in a cylindrical FAIMS device. *J.*
487 *Am. Mass Spectrom.* 25: 1274-12844
- 488 21. Shvartsburg AA (2009) Differential Ion Mobility Spectrometry; CRC Press: Boca Raton,
489 FL
- 490 22. Purves RW (2013) Enhancement of biological mass spectrometry by using separations
491 based on changes in ion mobility (FAIMS and DMS). *Anal Bioanal. Chem.* 405: 35-42.
- 492 23. Buryakov IA, Krylov EV, Nazarov EG, Rasulev UK (1993) A new method of separation
493 of multi-atomic ions by mobility at atmospheric pressure using a high-frequency amplitude-
494 asymmetric strong electric field. *Int. J. Mass Spectrom. Ion Proces.* 128: 143–148
- 495 24. Gabryelski W, Wu F, Froese KL (2008) Comparison of High-Field Asymmetric
496 Waveform Ion Mobility Spectrometry with GC Methods in Analysis of Haloacetic Acids in
497 Drinking Water. *Anal. Chem.* 75: 2478-2486
- 498 25. Ells B, Barnett DA, Purves RW, Guevremont R (2000) Detection of nine chlorinated and
499 brominated haloacetic acids at part-per-trillion levels using ESI-FAIMS-MS. *Anal. Chem.* 72,
500 5455-4559
- 501 26. Guevremont R, Purves RW, Barnett DA, Ells B (2006) FAIMS apparatus and method
502 using carrier gases that contain a trace amount of a dopant species. US Patent no. 7,026,612

- 503 27. Rorrer LCIII, Yost RA (2011) Solvent vapour effects on planar high-field asymmetric
504 waveform ion mobility spectrometry. *Int. J. Mass Spectrom.* 300: 173-181
- 505 28. Reeves K, Thomas K, Quilliam MA (2004) Mussel Tissue Certified Reference material
506 for Paralytic Shellfish Poisoning Toxins. Proceedings of the 5th International Conference on
507 Molluscan Shellfish Safety. Marine Institute, Pg. 116 - 122 (<http://hdl.handle.net/10793/576>,
508 Accessed May 23, 2014)
- 509 29. Dorr FA, Kovacevic B, Maksic ZB, Pinto E, Volmer DA (2011) Intriguing Differences in
510 the Gas-Phase Dissociation Behavior of Protonated and Deprotonated Gonyautoxin Epimers *J.*
511 *Am. Soc. Mass Spectrom.* 22: 2011-2020
- 512 30. Purves RW, Guevremont R (1999) Electrospray Ionization High-Field Asymmetric
513 Waveform Ion Mobility Spectrometry - Mass Spectrometry, *Anal. Chem.* 71: 2346-2357
- 514 31. Hall AB, Coy SL, Kafle A, Glick J, Nazarov E, Vouros P (2013) Extending the Dynamic
515 Range of the Ion Trap by Differential Mobility Filtration. *J. Am. Soc. Mass Spectrom.* 24: 1428-
516 1436.