NRC Publications Archive Archives des publications du CNRC

Modelling of Non-Linear Arterial Deformation and Friction Work during PTCA

Laroche, Denis; Delorme, Sebastien; Anderson, Todd; DiRaddo, Robert

NRC Publications Archive Record / Notice des Archives des publications du CNRC :
<https://nrc-publications.canada.ca/eng/view/object/?id=1cbba33c-17ea-4a1e-8caf-6975d0ceb065>
<https://publications-cnrc.canada.ca/fra/voir/objet/?id=1cbba33c-17ea-4a1e-8caf-6975d0ceb065>

Access and use of this website and the material on it are subject to the Terms and Conditions set forth at
<https://nrc-publications.canada.ca/eng/copyright>

READ THESE TERMS AND CONDITIONS CAREFULLY BEFORE USING THIS WEBSITE.

L'accès à ce site Web et l'utilisation de son contenu sont assujettis aux conditions présentées dans le site
<https://publications-cnrc.canada.ca/fra/droits>

LISEZ CES CONDITIONS ATTENTIVEMENT AVANT D'UTILISER CE SITE WEB.

Questions? Contact the NRC Publications Archive team at
PublicationsArchive-ArchivesPublications@nrc-cnrc.gc.ca. If you wish to email the authors directly, please see the first page of the publication for their contact information.

Vous avez des questions? Nous pouvons vous aider. Pour communiquer directement avec un auteur, consultez la première page de la revue dans laquelle son article a été publié afin de trouver ses coordonnées. Si vous n'arrivez pas à les repérer, communiquez avec nous à PublicationsArchive-ArchivesPublications@nrc-cnrc.gc.ca.

Modelling of Non-Linear Arterial Deformation and Friction Work during PTCA

Denis Laroche^a, Sebastien Delorme^a, Todd Anderson^b, and Robert DiRaddo^a

^aIndustrial Materials Institute (IMI), Boucherville, QC, Canada

^bUniversity of Calgary, Calgary, AB, Canada

Introduction

Percutaneous Transcatheter Coronary Angioplasty (PTCA) is the most common intervention for the treatment of a stenosed artery. Optimal results and minimal complications depend on the intervention strategy (balloon/stent size, inflation pressure) which is typically determined by angiographic images, patient clinical information and clinician's experience.

This work presents improvements to numerical technique for predicting the device/artery behavior during angioplasty. The goal of this numerical tool is to assist clinician in the selection of appropriate intervention strategy for a specific patient by using intravascular ultrasound (IVUS) imaging data to simulate the intervention.

Methods

A finite element modeling software developed at IMI for the analysis of large deformations of soft materials was used to solve angioplasty mechanics (Laroche 2003, Delorme 2004). The model computes the device/artery interaction and large deformations that occur during device insertion and deployment into the diseased artery. It predicts the resulting artery lumen patency, the stress distribution and the friction work distribution in the arterial wall.

The angioplasty balloon is modeled with membrane elements and the artery with incompressible solid elements, both with a hyperelastic constitutive model. An important contribution of this work is the development of a multi-body contact algorithm that is rapid and robust enough for handling complex contact and friction behavior between the balloon and the artery. The proposed algorithm performs implicit contact detection between virtual surfaces.

Results and conclusion

A proof-of-concept test was performed using IVUS images of one patient who underwent balloon angioplasty. A 3-D model of a 68 mm LAD artery segment was produced from IVUS images. The deployment of a 3.5 by 28 mm balloon was simulated by applying a 12 atm pressure.

The proposed model is capable of predicting the effect of intervention strategy on the artery deformation.

References

Delorme, S. et al (2004) *Annual Technical Conference (ANTEC) of the Society of Plastic Engineers*, Chicago, IL.
Laroche, D. et al (2003) *ASM Materials & Processes for Medical Devices Conference*, Anaheim, CA.